

BTICINO &gt; Wiring devices &gt; Solutions for healthcare establishment &gt; Call system



## CMCO3120

Plug for diagnostic contact - Mâtix. To be used with the socket for diagnostic contact. It can be used to disable the automatic alarm, when no automatic diagnostic device is connected. 1 module anthracite

### Technical features

Brand	BTicino
Standard reference	VDE0834-1-2 (DIN 41050)
Rated voltage	24Vac
Rated current	6A
N° of modules	1

### Related products



#### L4291

Socket mignon for auxiliary circuits 2P 6A  
24V SELV - tulip contacts d=3mm - centre distance 13mm - ...

We, Bticino S.p.A Viale Borri 231 21100 Varese (Italy), declare that all items listed in Bticino catalogues, have been manufactured in compliance with the principal elements of safety objectives of European Directive said LVD: 2014/35/EU: 26 February 2014 and, where requested, also in compliance with essential protection requirements of electromagnetic compatibility according to European Directive 2014/30/EU: 26 February 2014, and/or where requested also in compliance with 1995/5/CE: 9 March 1999 "R&TTE" or where requested also in compliance with 2014/53/EU: 16 April 2014 "RED". Bticino S.p.A. products are in compliance with the standard published by the International Electrotechnical Commission (IEC). The compliance can be proved by Certificates issued by organizations recognized by IEC according to the CB-scheme. Our items comply with relevant European Product- Standards and show, whether provided, CE marking, they have been constructed in accordance with good engineering practice in safety matters in force in the Community, they do not endanger the safety of persons, domestic animals or property when properly installed and maintained and used in applications for which they were made.